

## UnoVue™ Mouse/Rabbit HRP Detection Reagent

**Catalog No.** MRU-HRP100, MRU-HRP1000

**Intended Use** UnoVue Mouse/Rabbit HRP Detection Reagent is a non-biotin one-step detection reagent suitable for demonstrating antigens in formalin-fixed paraffin-embedded tissues and frozen sections. The UnoVue HRP Detection Reagent may also be used with blood smears, cytospreads, and cell preparations.

UnoVue HRP detection reagents have been developed by directly labeling anti-mouse and anti-rabbit immunoglobulins with enzymes using a proprietary tandem hyperlabelling technology. This ensures consistent and reproducible immunodetection of mouse and rabbit antibodies with a single reagent. Nuclear, cytoplasmic and membrane antigens in different types of tissues can be detected readily. The single step UnoVue Detection Reagent enables faster staining procedures than traditional two-step methods using biotin and avidin/streptavidin conjugates, with significantly lower background.

The UnoVue Mouse/Rabbit HRP Detection Reagent is suitable for use with all mouse and rabbit antibodies, both monoclonal and polyclonal. The reagents can be used for manual staining or with automated staining platforms.

| <b>Kit Contents</b> | <b>Description</b>                          | <b>Catalog #</b> | <b>Volume</b>       |
|---------------------|---|------------------|---------------------|
|                     | - UnoVue Mouse/Rabbit HRP Detection Reagent | MRU-HRP100       | 10 ml (100 Tests)   |
|                     |   | MRU-HRP1000      | 100 ml (1000 Tests) |

**Storage** Store at 2°-8°C. Do not freeze.

**Stability** 12-24 months (see expiration date on reagent bottles)

**Composition** UnoVue Mouse/Rabbit HRP ready to use reagent is formulated without azide or thimerosal preservatives.

**Material Required But Not Provided**

1. Xylene or dewaxing reagents
2. Absolute ethanol
3. Distilled or deionized water
4. Immuno Wash Buffer (DBS Cat# K005)
5. Pre-Blocking solution (DBS Cat# K023, optional)
6. Primary Antibody Diluent (DBS Cat# K004)
7. Counterstain
8. Mounting medium (DBS Cat# K002)
9. Chromogens (DBS cat# K047 Stable DAB Plus or Cat# K060 PermaYellow)
10. Tissue Primer or Peroxide Block (DBS Cat# K054/ K033)

**IVD: For In Vitro Diagnostic Use**

**DBS**

6166 Owens Drive, Pleasanton, CA 94588 Tel: 925 484 3350, Fax: 925 484 3390

Website: [www.dbiosys.com](http://www.dbiosys.com) e-mail: [customersupport@dbiosys.com](mailto:customersupport@dbiosys.com)

Page 1 of 2

- Precautions**
- i) Wear appropriate personal protective apparel. Avoid contact with clothes and exposed skin. In case of accidental skin exposure, flush with water immediately. Consult a physician if required.
  - ii) Interpretation of the results is the sole responsibility of the user.

- Recommended Staining Protocol**
1. Paraffin embedded tissue sections must be deparaffinized with xylene or dewaxing agent and rehydrated with a graded series of ethanol and water washes before staining. Follow the standard dewaxing and rehydration protocol used in your lab.
  2. The investigator needs to optimize the dilution and incubation times for primary antibodies.
  3. Each immunostaining run should include known positive and negative controls to assure proper functioning of the staining system and aid in valid interpretation of the results.

Typical controls:

Positive Control: A tissue known to contain the desired antigen which has yielded positive staining in the past.

Negative Controls:

Reagent Controls

- A. Substitute normal non-immune serum from the same host animal as the primary antibody (e.g. if using mouse monoclonal primary antibodies, use mouse non-immune serum).
- B. Substitute matching host species isotype control for primary antibody
- C. Use antigen-adsorbed primary antibody (i.e. antibody reagent which has been adsorbed with the target antigen to remove specific antibody)

Tissue control – A tissue known to *not* contain the desired antigen.

4. Consult the primary antibody supplier for recommended antigen recovery treatments. Perform epitope recovery pretreatments before starting the staining procedure.
5. Once the slide treatment has been started, DO NOT let tissues or specimens dry. This can cause undesirable background or artifacts.
6. After the primary antibody application, wash slides with Immuno Wash buffer (DBS cat# K 005).
7. Apply UnoVue Mouse/Rabbit HRP Detection Reagent (Make sure tissue sections/smears are completely covered with the reagent) and incubate for 20 min at room temperature.
8. Wash slides with Immuno Wash buffer then proceed to chromogen labeling and counterstain steps.

IVD: For In Vitro Diagnostic Use

**DBS**

6166 Owens Drive, Pleasanton, CA 94588 Tel: 925 484 3350, Fax: 925 484 3390

Website: [www.dbiosys.com](http://www.dbiosys.com) e-mail: [customersupport@dbiosys.com](mailto:customersupport@dbiosys.com)

Page 2 of 2